

The use of lubricants with the ThinPrep Pap test is not recommended. However, if a lubricant is necessary, consult this list of common lubricant brands for compatibility.<sup>1\*</sup>

# ThinPrep® Pap Test Lubricant Compatibility List

	Lubricant	Manufacturer	Contains Carbomer?
✓ Preferred <sup>†</sup>	<b>PAP Test Lubricating Jelly</b>	Aseptic Control Products	No
	<b>Surgilube Surgical Lubricant</b>	HR Pharmaceuticals	No
✓ Compatible <sup>‡</sup>	<b>K-Y Jelly (Physician Formula)</b>	Johnson & Johnson	No <sup>¶</sup>
	<b>Surgel</b>	Ulmer Pharmacal	No
✗ Incompatible <sup>§</sup>	<b>Aquagel Lubricating Gel</b>	Parker Laboratories, Inc.	Yes
	<b>Astroglide (Physician Formula)</b>	BioFilm, Inc.	No
	<b>Astroglide (Personal Formula)<sup>¶</sup></b>	BioFilm, Inc.	No
	<b>Crystelle</b>	Deltex Pharmaceuticals, Inc. - FDA injunction in 2011; no longer in operation	Unknown
	<b>HR Lubricating Jelly</b>	HR Pharmaceuticals, Inc.	Yes
	<b>Lubricating Gel</b>	Henry Schein	Yes
	<b>Lubricating Jelly</b>	McKesson	Yes
	<b>MediChoice Lubricating Jelly</b>	Owens & Minor	Yes
	<b>PDI Lubricating Jelly I and II</b>	PDI Healthcare	Yes
	<b>PSS Select (also known as Triad)</b>	PSS World Medical, Inc.	Yes
	<b>Rite Aid Pharmacy Lubricating Gel<sup>#</sup></b>	Rite Aid Corp.	Yes
	<b>Allegiance</b>	Medline Industries, Inc. (formerly Triad/H&P Industries)	Yes
	<b>Aplicare Sterile Lubricating Jelly (also known as Operand Lubricating Jelly)</b>	Aplicare Inc./Clorox Professional	Yes
	<b>Aqua Lube Personal Lubricant<sup>**</sup></b>	Mayer Laboratories	No
	<b>DynaLube Lubricating Jelly</b>	Dynarex Corporation	Yes
	<b>E-Z Lubricating Jelly</b>	Chester Packaging	Yes
	<b>IMCO Lubricating Jelly</b>	Medline Industries, Inc. (formerly Triad/H&P Industries)	Yes
	<b>Lubricating Jelly</b>	DUKAL Corporation	Yes
	<b>Lubri-Gel</b>	Sheffield Pharmaceuticals	Yes
	<b>Maxilube Personal Lubricant</b>	Mission Pharmacal	Yes
<b>NovaPlus</b>	Medline Industries, Inc. (formerly Triad/H&P Industries)	Yes	
<b>Pro Advantage Lubricating Jelly</b>	National Distribution & Contracting, Inc.	Yes	
<b>ReliaMed Lubricating Jelly</b>	ReliaMed	Yes	

\*The use of lubricants (e.g., KY Jelly) should be avoided prior to specimen collection. Lubricants can adhere to the filter membrane and may cause poor cell transfer to the slide. If its use is unavoidable, the lubricant should be used in minimum amounts. †Preferred: Lubricants have multiple lots run through periodic testing to ensure compatibility. ‡Compatible: Lubricants have been tested at least once, but periodic testing is not regularly performed. §Incompatible: Lubricants have either been tested and deemed incompatible or excluded from testing because they contain carbomer. ¶Carbomethylcellulose ingredient does not perform like carbomer-containing lubricants. ¶Not recommended due to low viscosity and cell reduction percentage. #Not recommended due to borderline cell reduction percentage. \*\*Not recommended due to low viscosity, personal lubricant.

1. ThinPrep 2000 System Operator's Manual. MAN-02585-001, Rev. 005. Marlborough, MA: Hologic, Inc.; 2014.

MISC-04037-001 Rev. 001 © 2016 Hologic, Inc. All rights reserved. Hologic, The Science of Sure and associated logos are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries. All other trademarks, registered trademarks, and product names are the property of their respective owners. The content in this piece is for information purposes only and is not intended to be medical advice. Information is intended for medical professionals in the U.S. and is not intended as a product solicitation or promotion where such activities are prohibited. Because Hologic materials are distributed through websites, eBroadcasts and tradeshow, it is not always possible to control where such materials appear. For specific information on what products are available for sale in a particular country, please contact your local Hologic representative or write to [diagnostic.solutions@hologic.com](mailto:diagnostic.solutions@hologic.com).